

OCT 29 2010

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130134</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>DEREK A MCGARRIGAN</u>	ID# <u>5420</u> ✓ Date <u>10/27/10</u> ✓
A Agency <u>Ketchikan Police Department</u>	Phone # <u>907-225-6631</u>
Instrument Location <u>Ketchikan Police Dept., 361 MAIN ST., Ketchikan, AK, 99901</u>	
B Alco S/N <u>70729 BMB</u> *	Target Value <u>.084</u> ✓ High Pressure <u>1100</u>
Alco Test Values <u>.088</u> ✓ <u>.087</u> ✓	
1st Alco 2nd Alco	
Signature <u>[Signature]</u>	
(OVER)	

(Do Not write in the area below)


\* Per Derck McGarrigan  
the alco serial number  
was 70729 not 130194.  
BMB  
1/5/11

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 5<sup>th</sup> day of JAN, 2011.



Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130134

BMB  
11/5/11

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130134

OCTOBER 27, 2010

OPERATOR'S NAME:

MCGARRIGAN DEREK A

OPERATOR'S NUMBER: 5420

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: 1234567

DEPT/AGENCY: KET1

CASE/REPORT: 10-2

TEST TYPE: V

ALCO TARGET VALUE: .004

ALCO S/N: 130194-70729

— BREATH ANALYSIS —

.004 ADJUSTED FOR 30.40 in  
ALCO TARGET .005 11:38  
BLANK TEST .000 11:39  
INTERNAL STANDARD VERIFIED 11:39  
ALCO TV 30.40 in .008 11:39  
BLANK TEST .000 11:40  
SUBJECT SAMPLE 11/5/11 .000 11:41  
BLANK TEST .000 11:41  
ALCO TV 30.40 in .007 11:42  
BLANK TEST .000 11:42

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130134

OCTOBER 27, 2010

TIME 11:35

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 30.40 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢